

Broadband Outdoor Radiometer Calibration Longwave

BORCAL-LW 2018-04

Calibration Facility

Southern Great Plains

Latitude: 36.605°N

Longitude: 97.488°W

Elevation: 317.0 meters AMSL

Time Zone: -6.0

Calibration date

06/06/2018 to 08/10/2018

Report Date

August 10, 2018

NOTICE

This report was prepared as an account of work sponsored by an agency of the United States government. Neither the United States government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States government or any agency thereof.

Broadband Outdoor Radiometer Calibration Report

Table of contents

Introduction.....	3
Control Instrument history plots.....	4
Results summary.....	5
Appendix 1 Instrument Details.....	A1-1
Appendix 2 BORCAL Notes.....	A2-1
Appendix 3 Session Configuration Audit Report.....	A3-1

Introduction

This report compiles the calibration results from a Broadband Outdoor Radiometer Calibration (BORCAL). The work was accomplished at the Radiometer Calibration Facility shown on the front of this report. The calibration results reported here are traceable to the World Infrared Standard Group (WISG).

This report includes these sections:

- Control Instruments - a group of instruments included in each BORCAL event that provides a measure of process consistency.
- Results Summary - a table of all instruments included in this report summarizing their calibration results and uncertainty.
- Instrument Details - the calibration certificates and application notes for each instrument.
- Environmental and Sky Conditions - meteorological conditions and reference irradiance during the calibration event.

Control Instrument History

Figure 1. Eppley PIR Control Instrument History (K0 Coefficient)




Figure 2. Eppley PIR Control Instrument History (K1 Coefficient)




Figure 3. Eppley PIR Control Instrument History (K2 Coefficient)





Figure 4. Eppley PIR Control Instrument History (K3 Coefficient)



Results Summary

Table 1. Results Summary

Instrument	Customer	K0 (W/m ²)	K1 (W/m ² /μV)	K2	K3	Kr * (K/μV)	U95 (W/m ²)	Page
29146F3	SGP	-7.9	0.25460	1.0369	-3.01	7.044e-4	±2.9	A1-2
29591F3	SGP	-12.1	0.22759	1.0402	-2.88	7.044e-4	±2.9	A1-5
30011F3	SGP	-13.6	0.27100	1.0410	-3.38	7.044e-4	±2.9	A1-8
30013F3	SGP	-16.0	0.27935	1.0467	-4.31	7.044e-4	±2.9	A1-11
30132F3	SGP	-8.2	0.23843	1.0324	-2.93	7.044e-4	±2.9	A1-14
30344F3	SGP	-11.9	0.23598	1.0406	-3.33	7.044e-4	±2.9	A1-17
30358F3	SGP	-12.9	0.23231	1.0453	-3.35	7.044e-4	±2.9	A1-20
30782F3	SGP	-7.4	0.23121	1.0322	-2.51	7.044e-4	±2.9	A1-23
30834F3	SGP	-13.7	0.25210	1.0468	-3.31	7.044e-4	±2.9	A1-26
30836F3	SGP	-9.8	0.24348	1.0344	-3.03	7.044e-4	±2.9	A1-29
36367F3	SGP	-13.1	0.31171	1.0427	1.04	7.044e-4	±3.1	A1-32
36368F3	SGP	-15.4	0.28008	1.0554	-3.65	7.044e-4	±2.9	A1-35

Note: Environmental Conditions for BORCAL starts on page A1-38.

* Kr used to derive K0,K1,K2, and K3

Appendix 1

Instrument Details

Calibration Certificates: 3 pages for each radiometer (4 including Environmental Conditions)

Environmental Conditions for BORCAL: Last Page of a Calibration Certificate. Note: This appears only once, at the end of Appendix 1.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 29146F3
Calibration Date: 8/10/2018 **Due Date:** 8/10/2019
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/6-15, 6/18-23, 6/26-30, 7/1-31, 8/1-10

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/19/2018	01/19/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 30835F3	05/08/2017	05/08/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31637F3	06/27/2017	06/27/2019

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Mike Dooraghi and Craig Webb

Michael Dooraghi, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Mike.Dooraghi@nrel.gov; 303-384-6329; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

29146F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-7.9
K_1	0.25460
K_2	1.0369
K_3	-3.01
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.25429
K_2	1.0205
K_3	-3.71
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.33
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.9

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.44
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.9

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 29591F3
Calibration Date: 8/10/2018 **Due Date:** 8/10/2019
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/6-15, 6/18-23, 6/26-30, 7/1-31, 8/1-10

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/19/2018	01/19/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 30835F3	05/08/2017	05/08/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31637F3	06/27/2017	06/27/2019

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Mike Dooraghi and Craig Webb

Michael Dooraghi, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Mike.Dooraghi@nrel.gov; 303-384-6329; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

29591F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-12.1
K_1	0.22759
K_2	1.0402
K_3	-2.88
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.22730
K_2	1.0129
K_3	-3.05
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.33
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 2.9

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.58
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30011F3
Calibration Date: 8/10/2018 **Due Date:** 8/10/2019
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/6-15, 6/18-23, 6/26-30, 7/1-31, 8/1-10

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/19/2018	01/19/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 30835F3	05/08/2017	05/08/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31637F3	06/27/2017	06/27/2019

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Mike Dooraghi and Craig Webb

Michael Dooraghi, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Mike.Dooraghi@nrel.gov; 303-384-6329; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30011F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 < 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 < 0$

K_0	-13.6
K_1	0.27100
K_2	1.0410
K_3	-3.38
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.27033
K_2	1.0101
K_3	-3.85
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 < 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.40
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 2.9

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.65
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30013F3
Calibration Date: 8/10/2018 **Due Date:** 8/10/2019
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/6-15, 6/18-23, 6/26-30, 7/1-31, 8/1-10

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/19/2018	01/19/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 30835F3	05/08/2017	05/08/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31637F3	06/27/2017	06/27/2019

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Mike Dooraghi and Craig Webb

Michael Dooraghi, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Mike.Dooraghi@nrel.gov; 303-384-6329; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30013F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-16.0
K_1	0.27935
K_2	1.0467
K_3	-4.31
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.27829
K_2	1.0106
K_3	-4.36
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.41
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 2.9

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.72
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30132F3
Calibration Date: 8/10/2018 **Due Date:** 8/10/2019
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/6-15, 6/18-23, 6/26-30, 7/1-31, 8/1-10

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/19/2018	01/19/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 30835F3	05/08/2017	05/08/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31637F3	06/27/2017	06/27/2019

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Mike Dooraghi and Craig Webb

Michael Dooraghi, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Mike.Dooraghi@nrel.gov; 303-384-6329; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30132F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-8.2
K_1	0.23843
K_2	1.0324
K_3	-2.93
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.23761
K_2	1.0141
K_3	-3.38
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.35
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.9

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.46
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.9

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30344F3
Calibration Date: 8/10/2018 **Due Date:** 8/10/2019
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/6-15, 6/18-23, 6/26-30, 7/1-31, 8/1-10

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/19/2018	01/19/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 30835F3	05/08/2017	05/08/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31637F3	06/27/2017	06/27/2019

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Mike Dooraghi and Craig Webb

Michael Dooraghi, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Mike.Dooraghi@nrel.gov; 303-384-6329; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30344F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-11.9
K_1	0.23598
K_2	1.0406
K_3	-3.33
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.23550
K_2	1.0140
K_3	-3.70
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.34
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 2.9

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.56
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30358F3
Calibration Date: 8/10/2018 **Due Date:** 8/10/2019
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/6-15, 6/18-23, 6/26-30, 7/1-31, 8/1-10

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/19/2018	01/19/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 30835F3	05/08/2017	05/08/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31637F3	06/27/2017	06/27/2019

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Mike Dooraghi and Craig Webb

Michael Dooraghi, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Mike.Dooraghi@nrel.gov; 303-384-6329; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30358F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-12.9
K_1	0.23231
K_2	1.0453
K_3	-3.35
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.23094
K_2	1.0160
K_3	-3.69
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.35
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.9

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.59
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30782F3
Calibration Date: 8/10/2018 **Due Date:** 8/10/2019
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/6-15, 6/18-23, 6/26-30, 7/1-31, 8/1-10

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/19/2018	01/19/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 30835F3	05/08/2017	05/08/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31637F3	06/27/2017	06/27/2019

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Mike Dooraghi and Craig Webb

Michael Dooraghi, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Mike.Dooraghi@nrel.gov; 303-384-6329; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30782F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-7.4
K_1	0.23121
K_2	1.0322
K_3	-2.51
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.23074
K_2	1.0161
K_3	-3.14
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.29
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 2.9

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.39
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 2.9

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30834F3
Calibration Date: 8/10/2018 **Due Date:** 8/10/2019
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/6-15, 6/18-23, 6/26-30, 7/1-31, 8/1-10

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/19/2018	01/19/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 30835F3	05/08/2017	05/08/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31637F3	06/27/2017	06/27/2019

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Mike Dooraghi and Craig Webb

Michael Dooraghi, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Mike.Dooraghi@nrel.gov; 303-384-6329; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30834F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-13.7
K_1	0.25210
K_2	1.0468
K_3	-3.31
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.25186
K_2	1.0166
K_3	-3.85
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.36
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 2.9

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.64
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30836F3
Calibration Date: 8/10/2018 **Due Date:** 8/10/2019
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/6-15, 6/18-23, 6/26-30, 7/1-31, 8/1-10

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/19/2018	01/19/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 30835F3	05/08/2017	05/08/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31637F3	06/27/2017	06/27/2019

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Mike Dooraghi and Craig Webb

Michael Dooraghi, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Mike.Dooraghi@nrel.gov; 303-384-6329; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30836F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-9.8
K_1	0.24348
K_2	1.0344
K_3	-3.03
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.24277
K_2	1.0121
K_3	-3.47
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.33
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 2.9

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.49
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 36367F3
Calibration Date: 8/10/2018 **Due Date:** 8/10/2019
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/6-15, 6/18-23, 6/26-30, 7/1-31, 8/1-10

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/19/2018	01/19/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 30835F3	05/08/2017	05/08/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31637F3	06/27/2017	06/27/2019

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Mike Dooraghi and Craig Webb

Michael Dooraghi, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Mike.Dooraghi@nrel.gov; 303-384-6329; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

36367F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-13.1
K_1	0.31171
K_2	1.0427
K_3	1.04
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.31122
K_2	1.0132
K_3	1.01
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.68
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.86
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.7
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.3

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 36368F3
Calibration Date: 8/10/2018 **Due Date:** 8/10/2019
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/6-15, 6/18-23, 6/26-30, 7/1-31, 8/1-10

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/19/2018	01/19/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 30835F3	05/08/2017	05/08/2019
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31637F3	06/27/2017	06/27/2019

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Mike Dooraghi and Craig Webb

Michael Dooraghi, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Mike.Dooraghi@nrel.gov; 303-384-6329; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

36368F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-15.4
K_1	0.28008
K_2	1.0554
K_3	-3.65
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.27902
K_2	1.0214
K_3	-4.45
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.41
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 2.9

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.4
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.68
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Environmental and Sky Conditions for BORCAL-LW 2018-04

Calibration Facility: Southern Great Plains

Latitude: 36.605°N


Longitude: 97.488°W

Elevation: 317.0 meters AMSL

Time Zone: -6.0

Reference Irradiance (hourly averages):

Figure 6. Reference Irradiance



Meteorological Observations (hourly averages):

Figure 7. Temperature




Figure 8. Humidity




Figure 9. Pressure




Figure 10. Estimated Precipitable Water Vapor (PWV)




Table 6. Meteorological Observations

Observations	Mean	Min	Max
Temperature (°C)	24.64	15.90	33.12
Humidity (%)	63.31	47.44	86.14
Pressure (mBar)	N/A	N/A	N/A
Est. Precipitable Water Vapor (mm)	32.8	18.7	52.5

For other information about the calibration facility visit: <http://www.arm.gov/docs/sites/sgp/sgp.html>

Appendix 2

BORCAL Notes

Instrument, Configuration, and Session Notes for the BORCAL

BORCAL Notes

Facility: Southern Great Plains

Comments:

Avg. Station Pressure and Temperature is for Tulsa, OK, which is used for the Solar Position Algorithm (SPA).

Session; Numbers: 10,16

Comments:

RCC: Recovered Session

Appendix 3

Session Configuration Audit Report

Latest Session Configuration Audit Report for the BORCAL

BORCAL/LW 2018-04 Session Configuration Audit Report

LOCATION

Facility	Facility Abbrev.	Contact	Latitude	Longitude	Elevation (m)	Avg press (mbr)	Avg temp (C)	Time zone	ISO
Southern Great Plains	SGP	Craig Webb	36.605	-97.488	317.0	992.0	15.0	-6.0	

SYSTEM

<p>% Error Thresholds</p> <p>TP(x) / TP(x-1) 25.0</p>	<p>Analysis Rejection</p> <p>Threshold 1 (Blue) 3.000</p> <p>Threshold 2 (Green) 4.000</p> <p>Threshold 3 (Brown) 5.000</p> <p>No. of Std. Dev. 3</p>	<p>Misc</p> <p>Scan Rate (s) 300</p> <p>Uncert. Significant Figures 2</p>
<p>Delta Thresholds</p> <p>Ref Pyg Stability 4.0</p> <p>Temp(x) - Temp(x-1) 5.0</p> <p>Hum(x) - Hum(x-1) 20.0</p> <p>Bar(x) - Bar(x-1) 5.0</p> <p>Thrm(x) - Temp(x) 10.0</p>	<p>Clock</p> <p>Reset Interval (m) 0</p> <p>Warning Threshold (s) 0</p> <p>Delta UT1 0.100</p>	<p>Auto Mode Zenith Angle</p> <p>Afternoon Startup 94</p> <p>Morning Shutdown 94</p>
		<p>Solar Position Algorithm</p> <p>Delta T (s) 69.084</p> <p>Atmos. Refraction (deg) 0.5667</p>

METEOROLOGICAL INSTRUMENTS

Channel	Junction Box	Cable	Location		
Temperature: E0710025T Vaisala HMP155 T					
59		36	36		
		Scale 100	Offset -40		
Humidity: E0710025H Vaisala HMP155 H					
63		18	18		
		Scale 100	Offset 0		
Pressure: None					
		Scale 0	Offset 0		
GPS TIME RECIEVER					
GPS: None					
Type	Port	Baud	Parity	Stop bits	Data bits
	0	0	0	0	0

DATALOGGER

Logger/Relay		DMM		Communications							
Unit 0	2014-1302 NREL RAP-DAQ		SG42000596 Agilent 34420A		Unit	Type	Addr.	Board	Parity	Stop	Data
Unit 0	None		None	DMM	0	GPIO	1	0	0	0	0
Unit 0	None		None	Relay	0	GPIO	4	1	0	0	0
Unit 0	None		None		-1		0	0	0	0	0
					-1		0	0	0	0	0
					-1		0	0	0	0	0
					-1		0	0	0	0	0
					-1		0	0	0	0	0
					-1		0	0	0	0	0
					-1		0	0	0	0	0
					-1		0	0	0	0	0
					-1		0	0	0	0	0
					-1		0	0	0	0	0
					-1		0	0	0	0	0
					-1		0	0	0	0	0
					-1		0	0	0	0	0
					-1		0	0	0	0	0
System Offsets:											
Volts DC (μ V)	0.43	0.00	0.00	0.00							
2-Wire Res. (mOhms)	2130.00	0.00	0.00	0.00							
4-Wire Res. (mOhms)	0.00	0.00	0.00	0.00							

BORCAL/LW 2018-04 Session Configuration Audit Report

PYRGEOMETER REFERENCE INSTRUMENTS

Cal Date	Cal Due Date	K0	Calibration Coefficients				Uncert. (W/m ²)	Max Out (mV)	Channel	Junction Box	Cable	Location	Active
			K1	K2	K3	Kr							
Pyrometer 1: 30835F3 Eppley PIR (Ventilated)													
05/08/2017	05/08/2019	-6.30900	0.23009	1.02060	-3.46670	7.04400E-4	1.80	9	5		2	T5-2	<input checked="" type="checkbox"/>
Pyrometer 1: Case 10K Temperature									4		2		
Pyrometer 1: Dome 10K Temperature									6		2		
Pyrometer 2: 31637F3 Eppley PIR (Ventilated)													
06/27/2017	06/27/2019	-19.35800	0.24987	1.06090	-3.43960	7.04400E-4	3.00	9	17		2	T6-2	<input checked="" type="checkbox"/>
Pyrometer 2: Case 10K Temperature									16		2		
Pyrometer 2: Dome 10K Temperature									18		2		

BORCAL/LW 2018-04 Session Configuration Audit Report

INSTRUMENTS

Serial Number / Model	Customer	Mfg RS	Ch	Box	Cable	Act	ISO	AIM	Sticker	Vent	Use	Kr	Location	Due
29146F3	SGP	3.6900	29		2	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T7-2	12
PIR	(Case 10K Temperature)		28		2									
	(Dome 10K Temperature)		30		2									
29591F3	SGP	4.1900	45		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T8-3	12
PIR	(Case 10K Temperature)		44		3									
	(Dome 10K Temperature)		46		3									
30011F3	SGP	3.5900	37		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T8-1	12
PIR	(Case 10K Temperature)		36		1									
	(Dome 10K Temperature)		38		1									
30013F3	SGP	3.5700	9		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T5-3	12
PIR	(Case 10K Temperature)		8		3									
	(Dome 10K Temperature)		10		3									
30132F3 ‡	SGP	3.9000	53		2	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T9-2	12
PIR	(Case 10K Temperature)		52		2									
	(Dome 10K Temperature)		54		2									
30344F3	SGP	3.9600	57		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T9-3	12
PIR	(Case 10K Temperature)		56		3									
	(Dome 10K Temperature)		58		3									
30358F3	SGP	4.2600	49		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T9-1	12
PIR	(Case 10K Temperature)		48		1									
	(Dome 10K Temperature)		50		1									
30782F3	SGP	4.0500	21		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T6-3	12
PIR	(Case 10K Temperature)		20		3									
	(Dome 10K Temperature)		22		3									
30834F3	SGP	3.7500	33		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T7-3	12
PIR	(Case 10K Temperature)		32		3									
	(Dome 10K Temperature)		34		3									
30836F3	SGP	3.9300	13		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T6-1	12
PIR	(Case 10K Temperature)		12		1									
	(Dome 10K Temperature)		14		1									
36367F3	SGP	3.0300	25		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T7-1	12
PIR	(Case 10K Temperature)		24		1									
	(Dome 10K Temperature)		26		1									
36368F3 ‡	SGP	3.0200	41		2	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T8-2	12
PIR	(Case 10K Temperature)		40		2									
	(Dome 10K Temperature)		42		2									

‡ Control Instrument