

Broadband Outdoor Radiometer Calibration Longwave

BORCAL-LW 2019-06

Generated by



Radiometer Calibration and Characterization

Calibration Facility Southern Great Plains

Latitude: 36.605°N
Longitude: 97.488°W
Elevation: 317.0 meters AMSL
Time Zone: -6.0

Calibration date
10/10/2019 to 11/18/2019

Report Date
November 20, 2019



NOTICE

This report was prepared as an account of work sponsored by an agency of the United States government. Neither the United States government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States government or any agency thereof.

Broadband Outdoor Radiometer Calibration Report

Table of contents

Introduction.....	3
Control Instrument history plots.....	4
Results summary.....	5
Appendix 1 Instrument Details.....	A1-1
Appendix 2 BORCAL Notes.....	A2-1
Appendix 3 Session Configuration Audit Report.....	A3-1

Introduction

This report compiles the calibration results from a Broadband Outdoor Radiometer Calibration (BORCAL). The work was accomplished at the Radiometer Calibration Facility shown on the front of this report. The calibration results reported here are traceable to the World Infrared Standard Group (WISG).

This report includes these sections:

- Control Instruments - a group of instruments included in each BORCAL event that provides a measure of process consistency.
- Results Summary - a table of all instruments included in this report summarizing their calibration results and uncertainty.
- Instrument Details - the calibration certificates for each instrument.
- Environmental and Sky Conditions - meteorological conditions and reference irradiance during the calibration event.

BORCAL Notes or Comments

PIR 32042F3 results were removed from this report due to bad data at the beginning of the calibration period.

PIR 32048F3 results were removed from this report due to intermittent bad case thermistor reporting throughout the calibration period.

Control Instrument History

Figure 1. Eppley PIR Control Instrument History (K0 Coefficient)




Figure 2. Eppley PIR Control Instrument History (K1 Coefficient)




Figure 3. Eppley PIR Control Instrument History (K2 Coefficient)





Figure 4. Eppley PIR Control Instrument History (K3 Coefficient)



Results Summary

Table 1. Results Summary

Instrument	Customer	K0 (W/m ²)	K1 (W/m ² /μV)	K2	K3	Kr * (K/μV)	U95 (W/m ²)	Page
29592F3	SGP	2.2	0.24065	1.0003	-3.54	7.044e-4	±3.0	A1-2
29596F3	SGP	1.2	0.22550	1.0025	-4.07	7.044e-4	±3.0	A1-5
30010F3	SGP	0.5	0.27129	1.0039	-4.01	7.044e-4	±3.1	A1-8
30014F3	SGP	0.5	0.24316	1.0007	-4.41	7.044e-4	±3.0	A1-11
30084F3	TWP	-2.3	0.25728	1.0044	-4.07	7.044e-4	±3.0	A1-14
30133F3	SGP	-5.1	0.23940	1.0129	-3.74	7.044e-4	±3.0	A1-17
30345F3	SGP	-2.3	0.26150	1.0065	-4.27	7.044e-4	±3.0	A1-20
30356F3	SGP	-1.1	0.24643	1.0040	-3.45	7.044e-4	±3.0	A1-23
30682F3	SGP	-1.2	0.23361	1.0027	-3.68	7.044e-4	±3.1	A1-26
30684F3	SGP	-2.7	0.24990	1.0084	-3.42	7.044e-4	±3.0	A1-29
30685F3	SGP	-0.7	0.23703	1.0023	-3.39	7.044e-4	±3.0	A1-32
30687F3	SGP	-0.3	0.23596	1.0007	-3.21	7.044e-4	±3.0	A1-35
30688F3	SGP	-1.0	0.23244	1.0002	-3.02	7.044e-4	±3.0	A1-38
30689F3	SGP	-2.3	0.25411	1.0024	-3.11	7.044e-4	±3.0	A1-41
30780F3	SGP	0.0	0.24349	1.0008	-3.35	7.044e-4	±3.0	A1-44
30784F3	SGP	-0.6	0.23893	1.0067	-3.52	7.044e-4	±3.0	A1-47
30835F3	SGP	1.0	0.22615	0.9958	-3.38	7.044e-4	±3.0	A1-50
30837F3	SGP	-1.2	0.24194	1.0027	-4.24	7.044e-4	±3.0	A1-53
31639F3	SGP	2.4	0.21886	0.9953	-3.77	7.044e-4	±3.0	A1-56
32041F3	SGP	1.7	0.24258	1.0009	-4.35	7.044e-4	±3.1	A1-59
32043F3	NSA	2.7	0.22126	0.9915	-3.07	7.044e-4	±3.0	A1-62
34304F3	AMF	-1.2	0.22784	0.9967	-4.25	7.044e-4	±3.0	A1-65
36280F3	AMF	1.2	0.31379	1.0027	-4.39	7.044e-4	±3.0	A1-68
36368F3	SGP	-0.8	0.27217	1.0099	-4.56	7.044e-4	±3.0	A1-71
37325F3	AMF	-0.6	0.26269	0.9985	-7.32	7.044e-4	±3.1	A1-74
37327F3	AMF	0.1	0.23224	0.9988	-3.92	7.044e-4	±3.0	A1-77
37332F3	AMF	0.4	0.24506	1.0016	-5.73	7.044e-4	±3.0	A1-80
37334F3	NSA	4.9	0.22056	0.9903	-6.04	7.044e-4	±3.0	A1-83
38865F3	SGP	7.1	0.30701	0.9875	-5.82	7.044e-4	±3.1	A1-86
38866F3	SGP	3.7	0.32907	0.9827	-6.40	7.044e-4	±3.0	A1-89
38867F3	SGP	0.9	0.29324	0.9917	-5.40	7.044e-4	±3.1	A1-92
38868F3	SGP	5.4	0.29589	0.9843	-6.11	7.044e-4	±3.0	A1-95
38869F3	SGP	4.9	0.29758	0.9871	-5.41	7.044e-4	±3.0	A1-98
38870F3	SGP	3.2	0.26446	0.9855	-4.69	7.044e-4	±3.0	A1-101

Note: Environmental Conditions for BORCAL starts on page A1-104.

* Kr used to derive K0,K1,K2, and K3

Appendix 1

Instrument Details

Calibration Certificates: 3 pages for each radiometer (4 including Environmental Conditions)

Environmental Conditions for BORCAL: Last Page of a Calibration Certificate. Note: This appears only once, at the end of Appendix 1.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 29592F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

29592F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	2.2
K_1	0.24065
K_2	1.0003
K_3	-3.54
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.24046
K_2	1.0061
K_3	-3.34
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.33
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.37
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrogeometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 29596F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

29596F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	1.2
K_1	0.22550
K_2	1.0025
K_3	-4.07
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.22536
K_2	1.0057
K_3	-3.97
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.37
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.38
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30010F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30010F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	0.5
K_1	0.27129
K_2	1.0039
K_3	-4.01
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.27135
K_2	1.0052
K_3	-3.97
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.44
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.44
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyregeometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30014F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30014F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	0.5
K_1	0.24316
K_2	1.0007
K_3	-4.41
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.24311
K_2	1.0020
K_3	-4.35
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.28
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.28
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30084F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: TWP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30084F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-2.3
K_1	0.25728
K_2	1.0044
K_3	-4.07
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.25772
K_2	0.9980
K_3	-4.00
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.20
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.27
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30133F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30133F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-5.1
K_1	0.23940
K_2	1.0129
K_3	-3.74
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.23944
K_2	0.9985
K_3	-3.55
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.24
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.41
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

[1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30345F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30345F3 Eppey PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-2.3
K_1	0.26150
K_2	1.0065
K_3	-4.27
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.26093
K_2	0.9996
K_3	-4.35
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.40
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.49
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30356F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30356F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-1.1
K_1	0.24643
K_2	1.0040
K_3	-3.45
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.24651
K_2	1.0009
K_3	-3.50
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.23
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.24
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30682F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30682F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-1.2
K_1	0.23361
K_2	1.0027
K_3	-3.68
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.23353
K_2	0.9994
K_3	-3.70
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.52
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.54
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30684F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30684F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-2.7
K_1	0.24990
K_2	1.0084
K_3	-3.42
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.24938
K_2	1.0005
K_3	-3.45
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.32
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.41
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30685F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30685F3 Eppey PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-0.7
K_1	0.23703
K_2	1.0023
K_3	-3.39
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.23714
K_2	1.0004
K_3	-3.40
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.24
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.24
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30687F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30687F3 Eppey PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-0.3
K_1	0.23596
K_2	1.0007
K_3	-3.21
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.23599
K_2	0.9998
K_3	-3.22
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.20
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.20
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30688F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30688F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-1.0
K_1	0.23244
K_2	1.0002
K_3	-3.02
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.23258
K_2	0.9974
K_3	-3.03
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.23
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.24
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30689F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30689F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-2.3
K_1	0.25411
K_2	1.0024
K_3	-3.11
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.25449
K_2	0.9962
K_3	-3.07
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.22
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.29
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrogeometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30780F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30780F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	0.0
K_1	0.24349
K_2	1.0008
K_3	-3.35
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.24349
K_2	1.0008
K_3	-3.35
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.23
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.23
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30784F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30784F3 Eppey PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-0.6
K_1	0.23893
K_2	1.0067
K_3	-3.52
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.23882
K_2	1.0049
K_3	-3.55
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.37
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.36
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30835F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30835F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,

V = thermopile output voltage (μV),

$W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),

where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),

where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,

$T_r = T_c + K_r * V$ = receiver temperature (K),

T_c = case temperature (K),

K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	1.0
K_1	0.22615
K_2	0.9958
K_3	-3.38
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.22601
K_2	0.9985
K_3	-3.34
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.23
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.25
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30837F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30837F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-1.2
K_1	0.24194
K_2	1.0027
K_3	-4.24
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.24212
K_2	0.9992
K_3	-4.25
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.24
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.26
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 31639F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

31639F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	2.4
K_1	0.21886
K_2	0.9953
K_3	-3.77
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.21855
K_2	1.0018
K_3	-3.57
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.22
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.29
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 32041F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

32041F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r)$$

[1]

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	1.7
K_1	0.24258
K_2	1.0009
K_3	-4.35
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.24210
K_2	1.0056
K_3	-4.19
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.44
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.45
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrogeometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 32043F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: NSA **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

32043F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	2.7
K_1	0.22126
K_2	0.9915
K_3	-3.07
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.22076
K_2	0.9988
K_3	-3.01
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.21
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.31
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 34304F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: AMF **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

34304F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-1.2
K_1	0.22784
K_2	0.9967
K_3	-4.25
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.22792
K_2	0.9934
K_3	-4.28
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.27
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.28
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 36280F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: AMF **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

36280F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	1.2
K_1	0.31379
K_2	1.0027
K_3	-4.39
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.31344
K_2	1.0058
K_3	-4.28
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.25
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.26
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyradiometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 36368F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

36368F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-0.8
K_1	0.27217
K_2	1.0099
K_3	-4.56
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.27239
K_2	1.0077
K_3	-4.60
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.25
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.26
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

[1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 37325F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: AMF **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

37325F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	-0.6
K_1	0.26269
K_2	0.9985
K_3	-7.32
K_r used to derive coefficients	7.044e-4

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.26280
K_2	0.9968
K_3	-7.37
K_r used to derive coefficients	7.044e-4

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.45
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.45
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 37327F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: AMF **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

37327F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	0.1
K_1	0.23224
K_2	0.9988
K_3	-3.92
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.23221
K_2	0.9992
K_3	-3.94
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.20
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.20
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 37332F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: AMF **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

37332F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	0.4
K_1	0.24506
K_2	1.0016
K_3	-5.73
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.24493
K_2	1.0027
K_3	-5.70
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.24
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.24
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 37334F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: NSA **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

37334F3 Eppey PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,

V = thermopile output voltage (μV),

$W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),

where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),

where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,

$T_r = T_c + K_r \cdot V$ = receiver temperature (K),

T_c = case temperature (K),

K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	4.9
K_1	0.22056
K_2	0.9903
K_3	-6.04
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.21899
K_2	1.0029
K_3	-5.15
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.32
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.48
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pygeometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 38865F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

38865F3 Eppey PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 \neq 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 \neq 0$

K_0	7.1
K_1	0.30701
K_2	0.9875
K_3	-5.82
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.30562
K_2	1.0052
K_3	-4.19
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 \neq 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.44
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.65
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.2

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 38866F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

38866F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	3.7
K_1	0.32907
K_2	0.9827
K_3	-6.40
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.32960
K_2	0.9935
K_3	-6.40
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.34
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.34
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyradiometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 38867F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

38867F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 \cdot V + K_2 \cdot W_r + K_3 \cdot (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,

V = thermopile output voltage (μV),

$W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),

where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),

where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,

$T_r = T_c + K_r \cdot V$ = receiver temperature (K),

T_c = case temperature (K),

K_r = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	0.9
K_1	0.29324
K_2	0.9917
K_3	-5.40
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.29311
K_2	0.9943
K_3	-5.37
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.43
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.45
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyradiometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 38868F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

38868F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r)$$

[1]

where,

K_0, K_1, K_2, K_3 = calibration coefficients,

V = thermopile output voltage (μV),

$W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),

where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),

where, $\sigma = 5.6704e-8 W \cdot m^{-2} \cdot K^{-4}$,

$T_r = T_c + K_r * V$ = receiver temperature (K),

T_c = case temperature (K),

K_r = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	5.4
K_1	0.29589
K_2	0.9843
K_3	-6.11
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.29418
K_2	0.9992
K_3	-5.04
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.35
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.54
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)



Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 38869F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

38869F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,
 $T_r = T_c + K_r * V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	4.9
K_1	0.29758
K_2	0.9871
K_3	-5.41
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.29656
K_2	1.0002
K_3	-4.79
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.33
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.50
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.6
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.1

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyradiometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 38870F3
Calibration Date: 11/18/2019 **Due Date:** 11/18/2020
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 10/10-18, 10/20-24, 10/26-29, 11/1-3, 11/5-18

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	01/10/2019	01/10/2020
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	01/10/2019	01/10/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	04/16/2018	04/16/2020
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	04/16/2018	04/16/2020

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

38870F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K_0 + K_1 * V + K_2 * W_r + K_3 * (W_d - W_r) \quad [1]$$

where,

K_0, K_1, K_2, K_3 = calibration coefficients,

V = thermopile output voltage (μV),

$W_d = \sigma * T_d^4$ = dome irradiance (W/m^2),

where, T_d = dome temperature (K),

$W_r = \sigma * T_r^4$ = receiver irradiance (W/m^2),

where, $\sigma = 5.6704e-8 \text{ W}\cdot\text{m}^{-2}\cdot\text{K}^{-4}$,

$T_r = T_c + K_r * V$ = receiver temperature (K),

T_c = case temperature (K),

K_r = efficiency coefficient (K/ μV).

Figure 1. Residuals for calc. vs ref. irradiance using $K_0 > 0$ Coefficients




Figure 2. Residuals for calc. vs ref. irradiance using $K_0 = 0$ Coefficients




Table 2. Calibration Coefficients for $K_0 > 0$

K_0	3.2
K_1	0.26446
K_2	0.9855
K_3	-4.69
K_r used to derive coefficients	$7.044e-4$

Table 3. Calibration Coefficients for $K_0 = 0$

K_0	0.0
K_1	0.26379
K_2	0.9942
K_3	-4.58
K_r used to derive coefficients	$7.044e-4$

Table 4. Uncertainty using $K_0 > 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.25
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Table 5. Uncertainty using $K_0 = 0$ Coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.5
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.36
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.5
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, U_{95} (W/m^2)	± 3.0

Figure 3. History of instrument (K0 Coefficient)




Figure 4. History of instrument (K1 Coefficient)




Figure 5. History of instrument (K2 Coefficient)





Figure 6. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyradiometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Environmental and Sky Conditions for BORCAL-LW 2019-06

Calibration Facility: Southern Great Plains

Latitude: 36.605°N


Longitude: 97.488°W

Elevation: 317.0 meters AMSL

Time Zone: -6.0

Page 4 of 4

Figure 6. Reference Irradiance



Meteorological Observations (hourly averages):

Figure 7. Temperature




Figure 8. Humidity




Figure 9. Pressure




Figure 10. Estimated Precipitable Water Vapor (PWV)




Table 6. Meteorological Observations

Observations	Mean	Min	Max
Temperature (°C)	6.04	-10.18	24.25
Humidity (%)	65.08	20.81	94.13
Pressure (mBar)	N/A	N/A	N/A
Est. Precipitable Water Vapor (mm)	11.0	3.5	23.0

For other information about the calibration facility visit: <http://www.arm.gov/docs/sites/sgp/sgp.html>

Appendix 2

BORCAL Notes

Instrument, Configuration, and Session Notes for the BORCAL

BORCAL Notes

Facility: Southern Great Plains

Comments:

Avg. Station Pressure and Temperature is for Tulsa, OK, which is used for the Solar Position Algorithm (SPA).

29596F3 Eppley PIR

Comments:

SGP address is 109596 Coal Road, Billings, OK 74630 tel 331-318-3354

Appendix 3

Session Configuration Audit Report

Latest Session Configuration Audit Report for the BORCAL

BORCAL/LW 2019-06 Session Configuration Audit Report

LOCATION									
Facility	Facility Abbrev.	Contact	Latitude	Longitude	Elevation (m)	Avg press (mbr)	Avg temp (C)	Time zone	ISO
Southern Great Plains	SGP	Craig Webb	36.605	-97.488	317.0	992.0	15.0	-6.0	

SYSTEM

% Error Thresholds TP(x) / TP(x-1) <input type="text" value="25.0"/>	Analysis Rejection Threshold 1 (Blue) <input type="text" value="3.000"/> Threshold 2 (Green) <input type="text" value="4.000"/> Threshold 3 (Brown) <input type="text" value="5.000"/> No. of Std. Dev. <input type="text" value="3"/>	Misc Scan Rate (s) <input type="text" value="300"/> Uncert. Significant Figures <input type="text" value="2"/>
Delta Thresholds Ref Pyg Stability <input type="text" value="4.0"/> Temp(x) - Temp(x-1) <input type="text" value="5.0"/> Hum(x) - Hum(x-1) <input type="text" value="20.0"/> Bar(x) - Bar(x-1) <input type="text" value="5.0"/> Thrm(x) - Temp(x) <input type="text" value="10.0"/>	Auto Mode Zenith Angle Afternoon Startup <input type="text" value="94"/> Morning Shutdown <input type="text" value="94"/>	Solar Position Algorithm Delta T (s) <input type="text" value="69.284"/> Atmos. Refraction (deg) <input type="text" value="0.5667"/>
Clock Reset Interval (m) <input type="text" value="0"/> Warning Threshold (s) <input type="text" value="0"/> Delta UT1 <input type="text" value="-0.100"/>		

METEOROLOGICAL INSTRUMENTS

Channel	Junction Box	Cable	Location
Temperature: E0710026T Vaisala HMP155 T			
<input type="text" value="239"/>			<input type="text" value="Temp"/>
		Scale <input type="text" value="100"/>	Offset <input type="text" value="-40"/>
Humidity: E0710026H Vaisala HMP155 H			
<input type="text" value="255"/>			<input type="text" value="Hum"/>
		Scale <input type="text" value="100"/>	Offset <input type="text" value="0"/>
Pressure: None			
		Scale <input type="text" value="0"/>	Offset <input type="text" value="0"/>

GPS TIME RECIEVER

GPS: None

Type	Port	Baud	Parity	Stop bits	Data bits
<input type="text" value="0"/>	<input type="text" value="0"/>	<input type="text" value="0"/>	<input type="text" value="0"/>	<input type="text" value="0"/>	<input type="text" value="0"/>

DATALOGGER

Logger/Relay		DMM		Communications						
Unit				Unit	Type	Addr.	Board	Parity	Stop	Data
Unit 0	2009-1206 NREL RAP-DAQ	MY42002863	Agilent 34420A	DMM	0	GPIB	21	0	0	0
Unit 1	2009-1207 NREL RAP-DAQ	MY42002864	Agilent 34420A	Relay	0	GPIB	24	1	0	0
Unit 2	2009-1208 NREL RAP-DAQ	MY42002866	Agilent 34420A	DMM	1	GPIB	22	0	0	0
Unit 3	2014-1302 NREL RAP-DAQ	SG42000596	Agilent 34420A	Relay	1	GPIB	25	1	0	0
				DMM	2	GPIB	23	0	0	0
				Relay	2	GPIB	26	1	0	0
				DMM	3	GPIB	1	0	0	0
				Relay	3	GPIB	4	1	0	0

	Unit 0	Unit 1	Unit 2	Unit 3
Cal Date	01/10/2019	01/10/2019	01/10/2019	01/10/2019
Cal Due Date	01/10/2020	01/10/2020	01/10/2020	01/10/2020
System Offsets: Volts DC (µV)	1.41	1.41	1.41	1.41
2-Wire Res. (mOhms)	2571.00	2571.00	2571.00	2571.00
4-Wire Res. (mOhms)	0.00	0.00	0.00	0.00

BORCAL/LW 2019-06 Session Configuration Audit Report

PYRGEOMETER REFERENCE INSTRUMENTS

Cal Date	Cal Due Date	Calibration Coefficients					Uncert. (W/m ²)	Max Out (mV)	Channel	Junction Box	Cable	Location	Active
		K0	K1	K2	K3	Kr							
Pyrometer 1: 31206F3 Eppley PIR (Ventilated)													
04/16/2018	04/16/2020	-0.20000	0.26400	0.99940	-3.26000	7.04400E-4	2.60	9	23		2	T5-2	<input checked="" type="checkbox"/>
Pyrometer 1: Case 10K Temperature									19		2		
Pyrometer 1: Dome 10K Temperature									27		2		
Pyrometer 2: 31237F3 Eppley PIR (Ventilated)													
04/16/2018	04/16/2020	3.50000	0.22892	0.99110	-3.69000	7.04400E-4	2.60	9	71		2	T6-2	<input checked="" type="checkbox"/>
Pyrometer 2: Case 10K Temperature									67		2		
Pyrometer 2: Dome 10K Temperature									75		2		

BORCAL/LW 2019-06 Session Configuration Audit Report

INSTRUMENTS

Serial Number / Model	Customer	Mfg RS	Ch	Box	Cable	Act	ISO	AIM	Sticker	Vent	Use	Kr	Location	Due
29592F3	SGP	3.7800	161		84	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	84/85	12
PIR	(Case 10K Temperature)		169		84									
	(Dome 10K Temperature)		170		85									
29596F3	SGP	4.2400	114		56	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	56/62	12
PIR	(Case 10K Temperature)		122		56									
	(Dome 10K Temperature)		41		62									
30010F3	SGP	3.2400	232		83	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	83	12
PIR	(Case 10K Temperature)		240		83									
	(Dome 10K Temperature)		248		83									
30014F3	SGP	3.7000	128		57	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	57/58	12
PIR	(Case 10K Temperature)		136		57									
	(Dome 10K Temperature)		137		58									
30084F3	TWP	3.5900	151		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T8-1	12
PIR	(Case 10K Temperature)		147		1									
	(Dome 10K Temperature)		155		1									
30133F3 ‡	SGP	3.9000	215		2	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T9-2	12
PIR	(Case 10K Temperature)		211		2									
	(Dome 10K Temperature)		219		2									
30345F3	SGP	3.7100	65		13	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	13/14	12
PIR	(Case 10K Temperature)		73		13									
	(Dome 10K Temperature)		74		14									
30356F3	SGP	3.8500	80		19	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	19/9	12
PIR	(Case 10K Temperature)		88		19									
	(Dome 10K Temperature)		10		9									
30682F3	SGP	3.9300	113		55	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	55/61	12
PIR	(Case 10K Temperature)		121		55									
	(Dome 10K Temperature)		40		61									
30684F3	SGP	3.7000	48		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	1/8	12
PIR	(Case 10K Temperature)		56		1									
	(Dome 10K Temperature)		9		8									
30685F3	SGP	3.7500	50		11	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	11/12	12
PIR	(Case 10K Temperature)		58		11									
	(Dome 10K Temperature)		72		12									
30687F3	SGP	3.7600	49		10	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	10/16	12
PIR	(Case 10K Temperature)		57		10									
	(Dome 10K Temperature)		16		16									
30688F3	SGP	3.9000	145		73	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	73/53	12
PIR	(Case 10K Temperature)		153		73									
	(Dome 10K Temperature)		33		53									
30689F3	SGP	3.4200	61		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	3/5/7	12
PIR	(Case 10K Temperature)		1		5									
	(Dome 10K Temperature)		8		7									
30780F3	SGP	3.7500	82		29	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	29/18	12
PIR	(Case 10K Temperature)		90		29									
	(Dome 10K Temperature)		18		18									
30784F3	SGP	3.7400	60		2	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	2/4/6	12
PIR	(Case 10K Temperature)		0		4									
	(Dome 10K Temperature)		2		6									

‡ Control Instrument

BORCAL/LW 2019-06 Session Configuration Audit Report

INSTRUMENTS

Serial Number / Model	Customer	Mfg RS	Ch	Box	Cable	Act	ISO	AIM	Sticker	Vent	Use	Kr	Location	Due
30835F3	SGP	4.0700	135		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T7-3	12
PIR	(Case 10K Temperature)		131		3									
	(Dome 10K Temperature)		139		3									
30837F3	SGP	3.8700	146		74	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	74/54	12
PIR	(Case 10K Temperature)		154		74									
	(Dome 10K Temperature)		34		54									
31639F3	SGP	4.2200	160		82	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	82/63	12
PIR	(Case 10K Temperature)		168		82									
	(Dome 10K Temperature)		42		63									
32041F3	SGP	3.7500	98		37	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	37/43	12
PIR	(Case 10K Temperature)		106		37									
	(Dome 10K Temperature)		24		43									
32042F3	SGP	3.8200	130		64	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	64/44	12
PIR	(Case 10K Temperature)		138		64									
	(Dome 10K Temperature)		25		44									
32043F3	NSA	4.1600	144		65	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	65/45	12
PIR	(Case 10K Temperature)		152		65									
	(Dome 10K Temperature)		26		45									
32048F3	SGP	3.7900	112		46	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	46/52	12
PIR	(Case 10K Temperature)		120		46									
	(Dome 10K Temperature)		32		52									
34304F3	AMF	3.9500	96		31	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	31/32	12
PIR	(Case 10K Temperature)		104		31									
	(Dome 10K Temperature)		105		32									
36280F3	AMF	3.0800	208		30	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	30	12
PIR	(Case 10K Temperature)		216		30									
	(Dome 10K Temperature)		224		30									
36368F3 ‡	SGP	3.0200	167		2	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T8-2	12
PIR	(Case 10K Temperature)		163		2									
	(Dome 10K Temperature)		171		2									
37325F3	AMF	3.7800	183		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T8-3	12
PIR	(Case 10K Temperature)		179		3									
	(Dome 10K Temperature)		187		3									
37327F3	AMF	3.9300	199		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T9-1	12
PIR	(Case 10K Temperature)		195		1									
	(Dome 10K Temperature)		203		1									
37332F3	AMF	3.7700	231		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T9-3	12
PIR	(Case 10K Temperature)		227		3									
	(Dome 10K Temperature)		235		3									
37334F3	NSA	4.3100	81		28	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	28/17	12
PIR	(Case 10K Temperature)		89		28									
	(Dome 10K Temperature)		17		17									
38865F3	SGP	2.8600	7		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T5-1	12
PIR	(Case 10K Temperature)		3		1									
	(Dome 10K Temperature)		11		1									
38866F3	SGP	2.6600	39		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T5-3	12
PIR	(Case 10K Temperature)		35		3									
	(Dome 10K Temperature)		43		3									

‡ Control Instrument

BORCAL/LW 2019-06 Session Configuration Audit Report

INSTRUMENTS

Serial Number / Model	Customer	Mfg RS	Ch	Box	Cable	Act	ISO	AIM	Stickr	Vent	Use	Kr	Location	Due
38867F3	SGP	2.9500	55		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T6-1	12
PIR	(Case 10K Temperature)		51		1									
	(Dome 10K Temperature)		59		1									
38868F3	SGP	2.9600	87		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T6-3	12
PIR	(Case 10K Temperature)		83		3									
	(Dome 10K Temperature)		91		3									
38869F3	SGP	2.9100	103		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T7-1	12
PIR	(Case 10K Temperature)		99		1									
	(Dome 10K Temperature)		107		1									
38870F3	SGP	3.3200	119		2	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T7-2	12
PIR	(Case 10K Temperature)		115		2									
	(Dome 10K Temperature)		123		2									