

Broadband Outdoor Radiometer Calibration Longwave

BORCAL-LW 2021-02

Generated by



Radiometer Calibration and Characterization

Calibration Facility

Southern Great Plains

Latitude: 36.605°N

Longitude: 97.488°W

Elevation: 317.0 meters AMSL

Time Zone: -6.0

Calibration date

06/23/2021 to 08/11/2021

Report Date

August 11, 2021



NOTICE

This report was prepared as an account of work sponsored by an agency of the United States government. Neither the United States government nor any agency thereof, nor any of their employees, makes any warranty, express or implied, or assumes any legal liability or responsibility for the accuracy, completeness, or usefulness of any information, apparatus, product, or process disclosed, or represents that its use would not infringe privately owned rights. Reference herein to any specific commercial product, process, or service by trade name, trademark, manufacturer, or otherwise does not necessarily constitute or imply its endorsement, recommendation, or favoring by the United States government or any agency thereof. The views and opinions of authors expressed herein do not necessarily state or reflect those of the United States government or any agency thereof.

Broadband Outdoor Radiometer Calibration Report

Table of contents

Introduction.....	3
Control Instrument history plots.....	4
Results summary.....	5
Appendix 1 Instrument Details.....	A1-1
Appendix 2 BORCAL Notes.....	A2-1
Appendix 3 Session Configuration Audit Report.....	A3-1

Introduction

This report compiles the calibration results from a Broadband Outdoor Radiometer Calibration (BORCAL). The work was accomplished at the Radiometer Calibration Facility shown on the front of this report. The calibration results reported here are traceable to the World Infrared Standard Group (WISG).

This report includes these sections:

- Control Instruments - a group of instruments included in each BORCAL event that provides a measure of process consistency.
- Results Summary - a table of all instruments included in this report summarizing their calibration results and uncertainty.
- Instrument Details - the calibration certificates for each instrument.
- Environmental and Sky Conditions - meteorological conditions and reference irradiance during the calibration event.

Control Instrument History

Figure 1. Eppley PIR Control Instrument (Residual means of current data using historical coefficients)




Figure 2. Eppley PIR Control Instrument History (K1 Coefficient)




Figure 3. Eppley PIR Control Instrument History (K2 Coefficient)





Figure 4. Eppley PIR Control Instrument History (K3 Coefficient)



Results Summary

Table 1. Results Summary

Instrument	Customer	K1 (W/m ² /μV)	K2	K3	Kr * (K/μV)	U95 (W/m ²)	Page
29148F3	SGP	0.24148	1.0026	-3.08	7.044e-4	±2.7	A1-2
29593F3	SGP	0.22685	1.0012	-3.58	7.044e-4	±2.7	A1-5
29595F3	SGP	0.21131	1.0053	-3.38	7.044e-4	±2.6	A1-8
30133F3	SGP	0.24106	1.0024	-3.06	7.044e-4	±2.7	A1-11
30682F3	SGP	0.23764	1.0012	-3.32	7.044e-4	±2.7	A1-14
30835F3	SGP	0.22723	0.9985	-3.33	7.044e-4	±2.6	A1-17
31391F3	TWP	0.24580	0.9972	-3.13	7.044e-4	±2.6	A1-20
32049F3	SGP	0.23574	1.0026	-3.26	7.044e-4	±2.6	A1-23
33057F3	TWP	0.25021	0.9971	-3.50	7.044e-4	±2.7	A1-26
33058F3	TWP	0.23867	0.9966	-3.03	7.044e-4	±2.6	A1-29
36368F3	SGP	0.27329	1.0087	-3.97	7.044e-4	±2.7	A1-32
37336F3	NSA	0.23325	0.9986	-3.83	7.044e-4	±2.6	A1-35
37339F3	NSA	0.21796	0.9941	-4.83	7.044e-4	±2.7	A1-38

Note: Environmental Conditions for BORCAL starts on page A1-41.

* Kr used to derive coefficients

Appendix 1

Instrument Details

Calibration Certificates: 3 pages for each radiometer (4 including Environmental Conditions)

Environmental Conditions for BORCAL: Last Page of a Calibration Certificate. Note: This appears only once, at the end of Appendix 1.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 29148F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

29148F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r)$$

[1]

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + Kr \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 Kr = efficiency coefficient (K/ μV).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.24148
K2	1.0026
K3	-3.08
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.43
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.4
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.7

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 29593F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

29593F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r)$$

[1]

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + Kr \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 Kr = efficiency coefficient (K/ μV).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.22685
K2	1.0012
K3	-3.58
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.40
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.4
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.7

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 29595F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

29595F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r) \tag{1}$$

[1]

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + Kr \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 Kr = efficiency coefficient (K/ μV).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.21131
K2	1.0053
K3	-3.38
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.36
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.3
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.6

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30133F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30133F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r) \tag{1}$$

[1]

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + Kr \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 Kr = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.24106
K2	1.0024
K3	-3.06
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.51
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.4
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.7

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30682F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30682F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r) \quad [1]$$

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + Kr \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 Kr = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.23764
K2	1.0012
K3	-3.32
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.42
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.4
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.7

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyregeometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 30835F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

30835F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r) \quad [1]$$

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + Kr \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 Kr = efficiency coefficient (K/ μV).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.22723
K2	0.9985
K3	-3.33
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.36
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.3
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.6

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyregeometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 31391F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: TWP **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

31391F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r) \quad [1]$$

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + Kr \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 Kr = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.24580
K2	0.9972
K3	-3.13
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.34
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.3
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.6

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 32049F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

32049F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r) \quad [1]$$

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + Kr \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 Kr = efficiency coefficient (K/ μV).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.23574
K2	1.0026
K3	-3.26
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.35
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.3
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.6

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 33057F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: TWP **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

33057F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r) \quad [1]$$

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + Kr \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 Kr = efficiency coefficient (K/ μV).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.25021
K2	0.9971
K3	-3.50
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.42
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.4
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.7

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 33058F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: TWP **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

33058F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r) \quad [1]$$

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + Kr \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 Kr = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.23867
K2	0.9966
K3	-3.03
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.37
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.4
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.6

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 36368F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: SGP **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

36368F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r) \quad [1]$$

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + Kr \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 Kr = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.27329
K2	1.0087
K3	-3.97
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.42
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.4
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.7

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

[1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 37336F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: NSA **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

37336F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r)$$

[1]

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + Kr \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 Kr = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.23325
K2	0.9986
K3	-3.83
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.37
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.4
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.6

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Southern Great Plains Radiometer Calibration Facility

National Renewable Energy Laboratory



Metrology Laboratory

Calibration Certificate

Test Instrument: Downwelling Pyrgeometer (Ventilated) **Manufacturer:** Eppley
Model: PIR **Serial Number:** 37339F3
Calibration Date: 8/11/2021 **Due Date:** 8/11/2023
Customer: NSA **Environmental Conditions:** see page 4
Test Dates: 6/23-25, 6/27-28, 6/30, 7/1-31, 8/1-11

This certifies that the above product was calibrated in compliance with procedure listed below. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the World Infrared Standard Group (WISG).

No statement of compliance with specifications is made or implied on this certificate. However, the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

Measurement Type	Instrument	Calibration Date	Calibration Due Date
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1206	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1207	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2009-1208	04/27/2021	04/27/2022
Data Acquisition	NREL Data Acquisition System Model RAP-DAQ, S/N 2014-1302	04/27/2021	04/27/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31206F3	02/19/2020	02/19/2022
Infrared Irradiance ‡	Eppley Downwelling Pyrgeometer Model PIR, S/N 31237F3	02/19/2020	02/19/2022

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure: SGP BORCAL-LW Calibration Procedure

Setup: Radiometers are calibrated outdoors, using the atmosphere as the source. Pyranometers and pyrgeometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits.

Calibrated by: Peter Gotseff and Craig Webb

Peter Gotseff, Technical Manager

Date

For questions or comments, please contact the technical manager at:

Peter.Gotseff@nrel.gov; 303-384-6327; 15013 Denver West Parkway, Golden, CO 80401, USA

Calibration Results

37339F3 Eppley PIR

The incoming irradiance (W_{in} , W/m^2) of the test instrument during calibration is calculated using this Measurement Equation:

$$W_{in} = K1 \cdot V + K2 \cdot W_r + K3 \cdot (W_d - W_r) \quad [1]$$

where,

$K1, K2, K3$ = calibration coefficients,
 V = thermopile output voltage (μV),
 $W_d = \sigma \cdot T_d^4$ = dome irradiance (W/m^2),
 where, T_d = dome temperature (K),

$W_r = \sigma \cdot T_r^4$ = receiver irradiance (W/m^2),
 where, $\sigma = 5.6704e-8 \text{ W} \cdot \text{m}^{-2} \cdot \text{K}^{-4}$,
 $T_r = T_c + K_r \cdot V$ = receiver temperature (K),
 T_c = case temperature (K),
 K_r = efficiency coefficient ($K/\mu V$).

Figure 1. Residuals for calculated using coefficients vs reference irradiance




Table 1. Calibration Coefficients

K1	0.21796
K2	0.9941
K3	-4.83
Kr used to derive coefficients	7.044e-4

Table 2. Uncertainty using coefficients

Type-B Standard Uncertainty, $u(B)$ (W/m^2)	± 1.3
Type-A Standard Uncertainty, $u(A)$ (W/m^2)	± 0.46
Combined Standard Uncertainty, $u(c)$ (W/m^2)	± 1.4
Effective degrees of freedom, $DF(c)$	+Inf
Coverage factor, k	1.96
Expanded Uncertainty, $U95$ (W/m^2)	± 2.7

Figure 2. History of instrument (Residual means of current data using historical BORCAL coefficients)




Figure 3. History of instrument (K1 Coefficient)




Figure 4. History of instrument (K2 Coefficient)





Figure 5. History of instrument (K3 Coefficient)



References:

- [1] Reda, I.; Stoffel, T. (2010). Pyrgometer Calibration for DOE-Atmospheric System Research Program using NREL Method (Presentation). 9 pp.; NREL Report No. PR-3B0-47756; <http://www.nrel.gov/docs/fy10osti/47756.pdf>.

Environmental and Sky Conditions for BORCAL-LW 2021-02

Calibration Facility: Southern Great Plains

Latitude: 36.605°N


Longitude: 97.488°W

Elevation: 317.0 meters AMSL

Time Zone: -6.0

Reference Irradiance (hourly averages):

Figure 6. Reference Irradiance



Meteorological Observations (hourly averages):

Figure 7. Temperature




Figure 8. Humidity




Figure 9. Pressure




Figure 10. Estimated Precipitable Water Vapor (PWV)




Table 6. Meteorological Observations

Observations	Mean	Min	Max
Temperature (°C)	23.58	15.88	31.65
Humidity (%)	89.33	62.25	100.04
Pressure (mBar)	N/A	N/A	N/A
Est. Precipitable Water Vapor (mm)	43.2	30.7	55.3

For other information about the calibration facility visit: <https://www.arm.gov/capabilities/observatories/sgp>

Appendix 2

BORCAL Notes

Instrument, Configuration, and Session Notes for the BORCAL

BORCAL Notes

Facility: Southern Great Plains

Comments:

Avg. Station Pressure and Temperature is for Tulsa, OK, which is used for the Solar Position Algorithm (SPA).

Appendix 3

Session Configuration Audit Report

Latest Session Configuration Audit Report for the BORCAL

BORCAL/LW 2021-02 Session Configuration Audit Report

LOCATION									
Facility	Facility Abbrev.	Contact	Latitude	Longitude	Elevation (m)	Avg press (mbr)	Avg temp (C)	Time zone	ISO
Southern Great Plains	SGP	Craig Webb	36.605	-97.488	317.0	992.0	15.0	-6.0	

SYSTEM

% Error Thresholds TP(x) / TP(x-1) <input type="text" value="25.0"/>	Analysis Rejection Threshold 1 (Blue) <input type="text" value="3.000"/> Threshold 2 (Green) <input type="text" value="4.000"/> Threshold 3 (Brown) <input type="text" value="5.000"/> No. of Std. Dev. <input type="text" value="3"/>	Misc Scan Rate (s) <input type="text" value="300"/> Uncert. Significant Figures <input type="text" value="2"/>
Delta Thresholds Ref Pyg Stability <input type="text" value="4.0"/> Temp(x) - Temp(x-1) <input type="text" value="5.0"/> Hum(x) - Hum(x-1) <input type="text" value="20.0"/> Bar(x) - Bar(x-1) <input type="text" value="5.0"/> Thrm(x) - Temp(x) <input type="text" value="10.0"/>	Auto Mode Zenith Angle Afternoon Startup <input type="text" value="94"/> Morning Shutdown <input type="text" value="94"/>	Solar Position Algorithm Delta T (s) <input type="text" value="69.384"/> Atmos. Refraction (deg) <input type="text" value="0.5667"/>
Clock Reset Interval (m) <input type="text" value="30"/> Warning Threshold (s) <input type="text" value="0"/> Delta UT1 <input type="text" value="-0.200"/>		

METEOROLOGICAL INSTRUMENTS

Channel	Junction Box	Cable	Location
Temperature: E0710026T Vaisala HMP155 T			
239			Temp
Scale		<input type="text" value="100"/>	Offset <input type="text" value="-40"/>
Humidity: E0710026H Vaisala HMP155 H			
255			Hum
Scale		<input type="text" value="100"/>	Offset <input type="text" value="0"/>
Pressure: None			
Scale		<input type="text" value="0"/>	Offset <input type="text" value="0"/>

GPS TIME RECIEVER

SGP Symmetricom NTP

Type	Port	Baud	Parity	Stop bits	Data bits
RS232	0	115200	0	1	8

DATALOGGER

Logger/Relay		DMM				Communications							
Unit						Unit	Type	Addr.	Board	Parity	Stop	Data	
Unit 0	2009-1206 NREL RAP-DAQ	MY42002863 Agilent 34420A				DMM	0	GPIB	21	0	0	0	0
Unit 1	2009-1207 NREL RAP-DAQ	MY42002864 Agilent 34420A				Relay	0	GPIB	24	1	0	0	0
Unit 2	2009-1208 NREL RAP-DAQ	MY42002866 Agilent 34420A				DMM	1	GPIB	22	0	0	0	0
Unit 3	2014-1302 NREL RAP-DAQ	SG42000596 Agilent 34420A				Relay	1	GPIB	25	1	0	0	0
		Unit 0	Unit 1	Unit 2	Unit 3	DMM	2	GPIB	23	0	0	0	0
Cal Date		04/27/2021	04/27/2021	04/27/2021	04/27/2021	Relay	2	GPIB	26	1	0	0	0
Cal Due Date		04/27/2022	04/27/2022	04/27/2022	04/27/2022	DMM	3	GPIB	1	0	0	0	0
System Offsets:	Volts DC (µV)	0.93	0.93	0.93	0.93	Relay	3	GPIB	4	1	0	0	0
	2-Wire Res. (mOhms)	2680.00	2680.00	2680.00	2680.00								
	4-Wire Res. (mOhms)	0.00	0.00	0.00	0.00								

BORCAL/LW 2021-02 Session Configuration Audit Report

PYRGEOMETER REFERENCE INSTRUMENTS

Cal Date	Cal Due Date	Calibration Coefficients					Uncert. (W/m ²)	Max Out (mV)	Channel	Junction Box	Cable	Location	Active
		K0	K1	K2	K3	Kr							
Pyrometer 1: 31206F3 Eppley PIR (Ventilated)													
02/19/2020	02/19/2022	0.00000	0.26576	0.99910	-3.67000	7.04400E-4	2.20	9	23		2	T5-2	<input checked="" type="checkbox"/>
Pyrometer 1: Case 10K Temperature									19		2		
Pyrometer 1: Dome 10K Temperature									27		2		
Pyrometer 2: 31237F3 Eppley PIR (Ventilated)													
02/19/2020	02/19/2022	3.80000	0.22978	0.99080	-4.34000	7.04400E-4	2.30	9	71		2	T6-2	<input checked="" type="checkbox"/>
Pyrometer 2: Case 10K Temperature									67		2		
Pyrometer 2: Dome 10K Temperature									75		2		

BORCAL/LW 2021-02 Session Configuration Audit Report

INSTRUMENTS

Serial Number / Model	Customer	Mfg RS	Ch	Box	Cable	Act	ISO	AIM	Sticker	Vent	Use	Kr	Location	Due
29148F3	SGP	3.6000	231		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T9-3	24
PIR	(Case 10K Temperature)		227		3									
	(Dome 10K Temperature)		235		3									
29593F3	SGP	4.1600	135		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T7-3	24
PIR	(Case 10K Temperature)		131		3									
	(Dome 10K Temperature)		139		3									
29595F3	SGP	4.4300	199		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T9-1	24
PIR	(Case 10K Temperature)		195		1									
	(Dome 10K Temperature)		203		1									
30133F3 ‡	SGP	3.9000	215		2	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T9-2	24
PIR	(Case 10K Temperature)		211		2									
	(Dome 10K Temperature)		219		2									
30682F3	SGP	3.9300	7		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T5-1	24
PIR	(Case 10K Temperature)		3		1									
	(Dome 10K Temperature)		11		1									
30835F3	SGP	4.0700	151		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T8-1	24
PIR	(Case 10K Temperature)		147		1									
	(Dome 10K Temperature)		155		1									
31391F3	TWP	3.7800	119		2	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T7-2	24
PIR	(Case 10K Temperature)		115		2									
	(Dome 10K Temperature)		123		2									
32049F3	SGP	3.8400	39		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T5-3	24
PIR	(Case 10K Temperature)		35		3									
	(Dome 10K Temperature)		43		3									
33057F3	TWP	3.8400	87		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T6-3	24
PIR	(Case 10K Temperature)		83		3									
	(Dome 10K Temperature)		91		3									
33058F3	TWP	3.8300	103		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T7-1	24
PIR	(Case 10K Temperature)		99		1									
	(Dome 10K Temperature)		107		1									
36368F3 ‡	SGP	3.0200	167		2	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T8-2	24
PIR	(Case 10K Temperature)		163		2									
	(Dome 10K Temperature)		171		2									
37336F3	NSA	3.9600	55		1	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T6-1	24
PIR	(Case 10K Temperature)		51		1									
	(Dome 10K Temperature)		59		1									
37339F3	NSA	4.4500	183		3	Yes	No	Yes	K0=0	Yes	PYG	7.044e-4	T8-3	24
PIR	(Case 10K Temperature)		179		3									
	(Dome 10K Temperature)		187		3									

‡ Control Instrument